

QUARTERLY QUALITY CONTROL CHECKS

2ND QUARTER 2015

LAB	ANALYST	Analyst's pH	Corr. Lab's pH	% Rel. Diff.	Analyst's CrVI	Corr. Lab's CrVI	% Rel. Diff.	Analyst's Tot.- Cr	Corr. Labs Tot. Cr	% Rel. Diff.	Analyst's Zn	Corr. Lab's Zn	% Rel. Diff.	Analyst's Ni	Corr. Labs Ni	% Rel. Diff.
MACI		6.99	6.86	2%	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A
		7.22	6.86	5%	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A
		7.27	6.86	6%	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A
Matthews Plating		8.81	8.8	0%	ND	<50	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A
JTF		6.78	6.71	1%	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A
		6.78	6.71	1%	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A
ELM #1		8.7	8.7	0%	0	0	0%	0.04	0.036	11%	0.67	0.676	1%	N/A	N/A	N/A
Repeat		9.6	9.6	0%	0.02	0	200%	0.06	0.036	50%	0.57	0.676	17%	N/A	N/A	N/A
		9.61	9.6	0%	0.02	0	200%	0.01	0.019	62%	0.19	0.208	9%	N/A	N/A	N/A
Great Lakes Metal Finishing		7.9	7.9	0%	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A
		7.8	7.9	1%	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A
Elm #2		6.92	7	1%	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A
		6.95	7	1%	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A
		6.94	7	1%	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A
Spring Arbor Coatings		8.31	7.7	8%	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A
		8.15	7.7	6%	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A
		8.25	7.7	7%	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A
		8.3	7.7	8%	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A
Classic Metal Finishing		7.9	8.61	9%	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A
		8.2	8.61	4%	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A
Eaton Aero- space- East		8	8.4	5%	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A
		8.13	8.4	3%	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A
		8.02	8.4	5%	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A
		8.03	8.4	5%	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A
		7.99	8.4	5%	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A
		8.19	8.4	3%	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A
		8.19	8.4	3%	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A
		8.02	8.4	5%	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A
2ND SHIFT		8.36	8.7	4%	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A
		7.9	8.4	6%	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A
		8.56	8.7	2%	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A

N/A--non-applicable

\*\*\*--low end of detection

Units are in ppb